




Sterile Processing 101: Cleaning, Disinfection, and Sterilization

Dr. Marjorie Wall, EDBA, CHL, CIS, CRCST, CSSBB



Objectives

- At the end of this lecture, participants will be able to:
 - Discuss facility, personnel and decontamination best practices for instrument reprocessing.
 - Explain instrument preparation, sterilization and sterile storage best practices.
 - Identify basic steps for high-level disinfection of flexible gastrointestinal endoscopes.



Standards

- In the United States, AAMI, AORN, ASHRAE and the CDC set the guidelines and best practices for instrument reprocessing.



Facility Design Ventilation

Functional Area	Airflow	Air Exchanges
Decontamination	Negative	10
Sterilizer Equipment Access	Negative	10
Sterilizer Loading and Unloading	Positive	10
Prep and Pack	Positive	10
Textile Pack Room	Positive	10
Clean/Sterile Storage	Positive	4

Facility Design Temperature/Humidity


Note: Conflicting Standards,
Recommend Conducting Risk
Assessment.

Functional Area	Temp	Humidity
Decontamination	60-65F	30-60%
General Work Area	68-73F	30-60%
Sterilization Equipment Room	75-85F	30-60%
Sterile Storage	Up to 75F	Up to 70%

Facility Design

- Clean and Dirty Areas must be physically separated or 36" apart.
- People flow must move from clean to dirty.
- Equipment flow must move from dirty to clean.

**No Cross
Contamination**



Environmental Consideration

- Floors and horizontal work surfaces
 - Clean and disinfect daily
 - Floor should be seamless, not grout
- Walls and storage shelves should be cleaned regularly on a scheduled basis
- Ceilings and Walls should not be made of shedding or porous materials



Sterile Processing Flow



Effective Cleaning CANNOT Take Place Without Effective Precleaning

- Why is that?
- Precleaning prevents formation of BIOFILM.
- Biofilm is “a group of microorganisms that form on a solid surface that comes in contact with water.”
- Biofilm can harbor resistant microorganisms reducing the effectiveness of sterilization.



Cleaning Starts on the Procedural Field

- Wipe instruments using a sterile, water moistened sponge.
- Instruments with lumens should be flushed with sterile water.
- Saline, bleach, or other solutions should NEVER be used.



Point of Use Cleaning



- Precleaning prevents damage of instrumentation and equipment.
- Dried blood is corrosive and causes pitting, rusting, and metal fatigue.
- Damaged instruments and equipment is unsafe to use on patients and can harbor microorganisms.



Biohazardous Transport

Must be Puncture Proof, Liquid Proof,
and Labeled as Biohazardous.



Decontamination PPE

SEQUENCE FOR PUTTING ON PERSONAL PROTECTIVE EQUIPMENT (PPE)

The type of PPE used will vary based on the level of precautions required, such as standard and contact, droplet or airborne infection isolation precautions. The procedure for putting on and removing PPE should be tailored to the specific type of PPE.

- 1. GOWN**
 - Fully cover torso from neck to knees, arms to end of wrists, and wrap around the back
 - Fasten in back of neck and waist
- 2. MASK OR RESPIRATOR**
 - Secure ties or elastic bands at middle of head and neck
 - Fit flexible band to nose bridge
 - Fit snug to face and below chin
 - Fit-check respirator
- 3. GOGGLES OR FACE SHIELD**
 - Place over face and eyes and adjust to fit
- 4. GLOVES**
 - Extend to cover wrist of isolation gown

USE SAFE WORK PRACTICES TO PROTECT YOURSELF AND LIMIT THE SPREAD OF CONTAMINATION

- Keep hands away from face
- Limit surfaces touched
- Change gloves when torn or heavily contaminated
- Perform hand hygiene




SEQUENCE FOR REMOVING PERSONAL PROTECTIVE EQUIPMENT (PPE)

Except for respirator, remove PPE at doorway or in anteroom. Remove respirator after leaving patient room and closing door.

- 1. GLOVES**
 - Outside of gloves is contaminated!
 - Grasp outside of glove with opposite gloved hand; peel off
 - Hold removed glove in gloved hand
 - Slide fingers of ungloved hand under remaining glove at wrist
 - Peel glove off over first glove
 - Discard gloves in waste container
- 2. GOGGLES OR FACE SHIELD**
 - Outside of goggles or face shield is contaminated!
 - To remove, handle by head band or ear pieces
 - Place in designated receptacle for reprocessing or in waste container
- 3. GOWN**
 - Gown front and sleeves are contaminated!
 - Unfasten ties
 - Pull away from neck and shoulders, touching inside of gown only
 - Turn gown inside out
 - Fold or roll into a bundle and discard
- 4. MASK OR RESPIRATOR**
 - Front of mask/respirator is contaminated — DO NOT TOUCH!
 - Grasp bottom, then top ties or elastics and remove
 - Discard in waste container

PERFORM HAND HYGIENE BETWEEN STEPS IF HANDS BECOME CONTAMINATED AND IMMEDIATELY AFTER REMOVING ALL PPE



REQUIRED PPE FOR DECONTAMINATION ROOM



DONNING:

1. Hair cover
2. Shoe covers
3. HH
4. Gown
5. Mask
6. Goggles or face shield
7. Gloves

DOFFING:

1. Shoe covers
2. Gown w/ gloves
3. Goggles or face shield
4. Mask
5. Hair cover
6. HH

Full PPE must be worn by decontamination staff or those performing decontamination tasks.

Others entering must don shoe covers, gown and hair cover.

Decontamination Chemicals

- Enzymatic?
- Validate Dosing Accurate
- Validate NOT Expired
- Validate Approved by OEM
- No Topping Off





AAMI ST108

Table 5—Frequency for water quality monitoring at water generation system

Water quality measurement	Type of testing	Routine monitoring sampling site	Minimum frequency of testing [*]	
			Utility Water	Critical Water
pH	pH meter** or Colorimetric dipsticks (sample tested within 15 minutes)	After the last treatment step	Quarterly	Monthly
Conductivity	Conductivity meter (in line or by measurement of a collected sample)	After the last treatment step, Storage tanks (if used)	Quarterly	Daily
Total Alkalinity	Colorimetric dipsticks Alkalinity test kit**	After the last treatment step, storage tanks (if used)	Quarterly	Monthly
Total Hardness	Determination of ppm as CaCO ₃ by Colorimetric dipsticks, Titration kit**, or Handheld meter**	After the last treatment step	Quarterly	Monthly
Bacteria	Heterotrophic plate count (see Annex H)	Loop out and loop return points	N/A	Monthly
Endotoxin	LAL test (see Annex H)	Loop out and loop return points	N/A	Monthly

^{*}NOTE 1 The recommendations for frequency of testing in this table are the recommended minimum frequency. If problems or issues arise with the water quality, it may be necessary to increase the frequency until they are resolved.

^{**}NOTE 2 When using these tests, the user should carefully follow the manufacturer's written IFU for accurate results. When measuring Critical Water levels, the water sample must be filled to the brim for these test kits, sealed and, if the sample is hot, allowed to cool to room temperature before testing. Testing within 15 minutes of sample collection prevents carbon dioxide absorption that can result in an inaccurate pH.

Decontamination Manual Cleaning

- Brushing of Instruments
- Nylon brushes are used to remove debris from instruments.
- Brushing should occur as follows:
 - Under water line to prevent aerosolization.
 - Brush all serrations
 - Brush all hinges.
 - Brush all lumens



Decontamination Ultrasonic Cleaning



Fig. 1 - Illustration of an imploding cavity in a liquid irradiated with ultrasound




Decontamination Automated Cleaning

- Place instruments in a position ensuring maximum exposed surface area through the automated wash process.
- Stringer should be placed so that hinged instruments are held in the open position.



Decontamination QA Testing





World Laparoscopy Hospital



Assembly Inspection

- Check each instrument for the following:
 - Corrosion
 - Rust
 - Pitting
 - Cracks
 - Burrs
 - Sign of wear
- If any of the above are found, remove the instrument from service.



Assembly Inspection

- Check each instrument for functionality:
- Scopes
 - Visual inspect lens for cracks or water penetration
 - Verify optics not damaged
- Cameras
 - Verify prisms not cracked or wet internally
 - Check for damaged cords
- Light Sources
 - Verify fiber optics not damaged



Assembly Stains

Stain Color	Probable Cause
Brown/Orange Stains	High pH - improper soaps, baked on blood, soaking in saline or using laundry soap (usually is not rust)
Bluish-Black Stains	Exposure to saline, blood or potassium chloride Reverse plating if two types of metals are placed in ultrasonic together
Light and Dark Spots	Water spots from allowing instrument to air dry
Dark Brown/Black Stains	Low pH acid stain - detergents or dried blood
Multi-Color Stains	Excessive heat - “hot spots” in autoclave
Bluish-Gray Stains	Cold sterilization solution used outside manufacturer guidelines

Assembly Chemical Integrators






Packaging Peel Pouches

- Size of Peel Pouch: 1" around instrument.
- Handle located by chevron
- Chemical integrator included
- Tip Protector if appropriate
- Double Peel Pack only if validated
- Do not overload with too many instruments
- Do not use inside trays
- Sterilize paper to plastic on side



Packaging Polypropylene Wrap



Packaging Rigid Containers



Assembly QA

- OR leadership documents tray defects in SPM.
- When process out of control, partner with SPD leadership to fix.

SPM Quality Event - Create - Details

Event ID: 12
Event type: Product
Event date: 10/11/2017
Status: Not reported

Event details (required):

Requirements:

Requirement	Pass	Fail	Not reviewed
Prepare/Package			
Includes an internal chemical indicator	Pass	Fail	Not reviewed
Has a count sheet	Pass	Fail	Not reviewed
Count sheet filled out completely	Pass	Fail	Not reviewed
Container fillers in place and intact	Pass	Fail	Not reviewed
Properly labeled	Pass	Fail	Not reviewed
Packaging sized appropriately	Pass	Fail	Not reviewed
Packaging sealed appropriately	Pass	Fail	Not reviewed
Packaging is free of holes and tears	Pass	Fail	Not reviewed
All instruments functional	Pass	Fail	Not reviewed
Sterilization			
Internal chemical indicator has passed	Pass	Fail	Not reviewed
Point of use			
All Instruments are the Correct Type and Size	Pass	Fail	Not reviewed
No Instruments are missing	Pass	Fail	Not reviewed
All Instruments are functional at the point of use	Pass	Fail	Not reviewed
Instrument Tray Contaminated	Pass	Fail	Not reviewed
Instruments have been pre-treated	Pass	Fail	Not reviewed

Event info: Copy info to details
Low Brainerd - 000002710, Assembly by Robert Zangola on 10/11/2017 3:43 PM (2017)
Responsible user: Lucknow

Previous Next

Point of use

All Instruments are the Correct Type and Size

Pass Fail Not reviewed

No Instruments are missing

Pass Fail Not reviewed

All Instruments are functional at the point of use

Pass Fail Not reviewed

Instrument Tray Contaminated

Pass Fail Not reviewed

Instruments have been pre-treated

Pass Fail Not reviewed

IUSS Sterilization

- Rapid Sterilization Process for Emergency Use
- Instruments Must Be Validated by OEM
- Implants Should Not Be IUSS'd
- Transport Closed Container
- Can Be Wet
- Tray Cannot be Stored for Another Patient



Prevac Steam Sterilization

- Minimum 270F 4min 20min Dry
- Must Not Be Wet
- Load Configuration:
 - Linen
 - Peel Pouches
 - Wrapped Items
 - Rigid Containers



Prevac Sterilization QA

- Required Testing:
 - Bowie Dick Test 1st Load Daily
 - Biological for Weakest Cycle Daily
 - Biological in Every Implant Load
- Sterilizer Qualification Testing
 - 3 Consecutive Biological
 - 3 Consecutive Bowie Dick
- Document Lot Number and Expiration Date
- Start a new control daily or when lot number changes






Low Temperature Sterilization

- H2O2 Based Technology
 - Plasma or Vaporized
- Used for Heat Sensitive Items
- Only Items Validated for Cycle Can be Ran
- Weight Restrictions on Loads
- No Porous or Absorbent Materials (paper)
- Only Chemical Integrators and Tape Validated can be used



Sterile Integrity

- Items should be stored as not to crush, compress, puncture or compromise sterility of contents.
- Whenever there is a question as to whether the package is sterile or not, it is considered unsterile.



Sterile Storage

- Configure so Wrapped Trays are Not Stacked
- Items are Stored Ergonomically Correct




Sterile Transport

- Sterile Trays should be covered during transport to protect from contamination.
- Trays should be handled minimally to prevent damage to packaging.




Robotic Inst Considerations

- Davinci Robotic Instruments require special processes to clean and sterilize.
- Routine Direct Observation Competencies
- Specialized Sonic
- Specialized Rigid Containers and Wraps
- Robotic Check for QA





TASS Considerations


- Multiple Outbreaks
- RCA completed
- Enzymatic Detergents
- Eye Instruments MUST go through a full rinse, preferably with deionized water.
- Lumens must be flushed with water profusely prior to sterilization



CJD Considerations

- CJD Risk for any procedures involving dura matter, spinal fluid, back of eye
- If unknown, treat as CJD.
- Process required for surgery to communicate to SPD
- Internal risk assessment for how to handle trays
- Single Use Instruments?
- CJD Cycle: 134C for 18 minutes








Flexible Scopes Standards

<https://www.cdc.gov/hicpac/recommendations/flexible-endoscope-reprocessing.html>

Standards of Infection Control in Reprocessing of Flexible Gastrointestinal Endoscopes

Standards of Infection Control in
Reprocessing of Flexible Gastrointestinal
Endoscopes





Flexible Scopes

Point of Use Cleaning

Flexible Scopes Processing Steps

- Follow the OEM IFU!
 - Pre-Clean Scope
 - Leak Test Scope
 - Manual Clean
 - Follow Directions Exactly.
 - Includes: Brushing, Flushing, and or Suctioning
 - Rinse after Cleaning
 - Visual Inspection
 - High Level Disinfection
 - Rinse after HLD
 - Dry (alcohol if required by OEM)
 - Store

Flexible Scope Decontamination

- Three Bay Sink
 - Sink 1 Leak Test
 - Sink 2 Soak/Brush/ Flush
 - Sink 3 Rinse
- Dirty to Clean Flow
- 36" or physical barrier between dirty and clean
- Same PPE and Chemical Requirements as SPD



Manual High Level Disinfection

- Ensure chemical is validated by Scope OEM
- Manual Solutions require specific time, temperature, and length of use following manufacturer IFU.



Manual High Level Disinfection

- Chemical may require activation.
- Date chemical opened and date expired after opening
- Chemical Strip Opening and Testing
- After expiration, chemical may need to be neutralized prior to disposal.
- PPE should always be worn when handling chemicals.
- Never Top Off
- Have a Spill Kit and Eye Wash



Manual High Level Disinfection

- Scope must stay fully submerged for OEM required time
- Temperature must meet OEM requirement
- Rinse must be completed with sterile or filtered water according to OEM



***CIDEX SOLUTIONS TRAY NOT INCLUDED**

Automated Endoscope Reprocessor HLD



- Ensure your scope is validated for your machine.
 - Ensure you have all required attachments
- Scope washers do not replace manual cleaning

Automated Endoscope Reprocessor HLD

- Machine should be self disinfecting
- No Residual water should remain in hoses and reservoirs
- Cycles for alcohol flushing and forced air drying are desirable
- Self Contained or external water filtration system
- Follow Scope OEM and AER OEM instructions exactly



High Level Disinfection Documentation



- GI Scope Processing must be traceable between patients
- Time between steps documentation is Best Practice (if too much time elapses, extended processing is required)
- Leak Testing pass or fail must be documented
 - If fail, scope must be sent out for repair
- Chemical Strip Test should be completed and documented daily and on each cycle.
- Temperature of Chemical Solution
- Exposure Time of Chemical Solution

Flexible Scopes Storage

- Storage cabinets should be made of a material that can be disinfected.
- In Conventional Storage (no drying options) scopes are hung vertically to facilitate drying.
- When using drying cabinets, follow OEM IFU for how to position scopes.



Flexible Scopes Storage

- Removable buttons and valves should be reprocessed and stored with the scope as a unique set for tracking patient to patient
- Hung scopes should not come in contact with each other, and gloves should always be worn to prevent cross contamination.
- **SGNA Supports a 7 Day Hang Time**
 - Do a Risk Assessment





Dr. Marjorie Wall

Former President HSPA Board of Directors



Dr. Marjorie Wall, EDBA, CSSBB

Associate Director, Sterile Processing

marjorie.wall@cshs.org

Cedars-Sinai

8700 Beverly Blvd., North Tower.5212 : Los Angeles CA 90048

office 310-423-3232 : mobile 714-316-4452 : cedars-sinai.org